elman-FURNACE PRIVATE LIMITED
MAKING YOUR FUTURE STRONGER
SF NO.380/2,Athiparasakthi Nagar,Seerapalayam,Madukkarai,Coimbatore - 641 105.
+91 9894103505, 9500964441, 9500964442
GST:33AAHCE7967J1Z5
elman@elmanfurnace.com,elmancalibration@gmail.com
www. elmanfurnace.com
Mfrs.All types of Furnace & Ovens, PWHT Machine,Recorders, Sensors, Thermocouples, Heaters, Control Instruments, Calibration and Services.

CALIBRATION CERTIFICATE

Certificate Number: ELE/25-26/4/1

Customer Name & Address	ACCENT STEELS INDIA P LTD S.F.No.234/1B Sakthi Main Road Kunnathur Village, Coimbatore - 641 107	**Date of Receipt**	06/04/2026
		Date of Calibration	06/04/2026
		Recommended Due Date	05/07/2026
		Date of Issue of Certificate	08/04/2026

DETAILS OF UNIT UNDER CALIBRATION (UUC)

Description	: TEMPERATURE CONTROLLER	**Make**	: GIC
Id No	: TC-42	**Resolution**	: 1 °C
Cal At	: Onsite	**Range**	: 0 to 1200 °C
Location of UUC	: HEATTREATMENT FURNACE	**Condition of UUC**	: Satisfactory
Type	: Cr/Al		

Environmental Conditions		**Calibration Method Used**	
Temperature	25.3 °C	**Calibration Procedure**	EE-SOP-04
Humidity	49 %RH	**Reference Standards**	EUROMET CG11

DETAILS OF STANDARD USED FOR CALIBRATION
Standards Used For Calibration Are Traceable To National Standard

Description	Id/Sr. No.	Certificate No	Validity
MULTIFUNCTION CALIBRATOR	ELE/CAL/004 / 22060988	RPL-2025-0847-C-007994	18/08/2026

Discipline : Electro-Technical

Calibration Results

Standard Reading	UUC Reading	Error	Expanded Uncertainty (±)
°C	°C	°C	°C
50.0	49	-1.0	0.67
300.0	299	-1.0	0.67
500.0	499	-1.0	0.67
800.0	799	-1.0	0.67
1000.0	999	-1.0	0.67
1200.0	1199	-1.0	0.67

Calibrated By		**Approved By**
Elman Furnace		**Elman Furnace**
Calibration Engineer		**Quality Manager**
Format No.:EE-F-7.8-01		**Issue No:02/25.03.2024, Rev:01/03.03.2025**

Results of this report only relate to the Item(s) mentioned and Calibrated by elman-FURNACE PRIVATE LIMITED
This certificate shall not be reproduced in full without written approval of elman-FURNACE PRIVATE LIMITED
All the readings noted here are valid as per the conditions and measurements at the time of calibration

- End of Calibration Certificate -